



February 20, 2025

Nevada Medicaid Web Announcement 3558

Oxygen Concentrator Recall: JIANGSU JUMAO X-CARE MEDICAL EQUIPMENT CO LTD Removes JMC5A Ni/TruAire-5 Oxygen Concentrator Due to the Devices Spontaneously Catching Fire

JIANGSU JUMAO X-CARE MEDICAL EQUIPMENT CO LTD is recalling devices due to incidents of device melting and fire during the use of oxygen concentrators. The cause of the fires and melting remains under investigation.

This [recall](#) involves removing certain devices from where they are used or sold. The U.S. Food and Drug Administration (FDA) has identified this recall as a Class I recall, the most serious type. This device may cause serious injury or death if you continue to use it.

See the [full recall notice](#) on the FDA's website for more information and the next actions to take.

Customers in the U.S. with questions about this recall should contact Compass Health Brands Corp at 1-800-376-7263 x444.